

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100923\
 Data File : VN079472.D
 Acq On : 09 Oct 2023 15:11
 Operator : JC\MD
 Sample : 04739-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW13

Quant Time: Oct 09 23:15:48 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100523W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 06 03:34:49 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	250515	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	507612	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	470292	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	141271	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	217285	53.493	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	106.980%
35) Dibromofluoromethane	8.171	113	187678	51.448	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.900%
50) Toluene-d8	10.571	98	590441	44.733	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	89.460%
62) 4-Bromofluorobenzene	12.853	95	189419	44.718	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.440%
Target Compounds						
30) Chloroform	7.971	83	107023	16.159	ug/l	92
47) Bromodichloromethane	9.894	83	12955	2.270	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100923\
Data File : VN079472.D
Acq On : 09 Oct 2023 15:11
Operator : JC\MD
Sample : 04739-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW13

Quant Time: Oct 09 23:15:48 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100523W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 06 03:34:49 2023
Response via : Initial Calibration

