

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100923\
 Data File : VN079473.D
 Acq On : 09 Oct 2023 15:36
 Operator : JC\MD
 Sample : 04740-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RW4

Quant Time: Oct 09 23:16:09 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100523W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 06 03:34:49 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	246467	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	505015	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	463481	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	139704	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	216539	54.185	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	108.380%
35) Dibromofluoromethane	8.171	113	185213	51.034	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.060%
50) Toluene-d8	10.571	98	582882	44.388	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	88.780%
62) 4-Bromofluorobenzene	12.853	95	185334	43.979	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	87.960%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100923\
Data File : VN079473.D
Acq On : 09 Oct 2023 15:36
Operator : JC\MD
Sample : 04740-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RW4

Quant Time: Oct 09 23:16:09 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100523W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 06 03:34:49 2023
Response via : Initial Calibration

