

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100923\
 Data File : VN079474.D
 Acq On : 09 Oct 2023 16:00
 Operator : JC\MD
 Sample : 04740-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BW1D

Quant Time: Oct 09 23:16:28 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100523W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 06 03:34:49 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	253063	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	516968	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	489961	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	156518	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	220129	53.648	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 107.300%	
35) Dibromofluoromethane	8.171	113	192541	51.826	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 103.660%	
50) Toluene-d8	10.571	98	621591	46.241	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 92.480%	
62) 4-Bromofluorobenzene	12.853	95	202931	47.041	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 94.080%	
Target Compounds						
					Qvalue	
16) Acetone	4.442	43	12981	7.976	ug/l #	77
40) Benzene	8.606	78	9994	0.604	ug/l	92
51) 4-Methyl-2-Pentanone	10.447	43	26174	4.770	ug/l	91
65) Chlorobenzene	11.894	112	203115	18.615	ug/l	99
91) 1,2-Dichlorobenzene	14.106	146	4970	0.900	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100923\
Data File : VN079474.D
Acq On : 09 Oct 2023 16:00
Operator : JC\MD
Sample : 04740-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BW1D

Quant Time: Oct 09 23:16:28 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100523W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 06 03:34:49 2023
Response via : Initial Calibration

