

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100923\
 Data File : VN079476.D
 Acq On : 09 Oct 2023 16:48
 Operator : JC\MD
 Sample : 04740-03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW3

Quant Time: Oct 09 23:17:35 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100523W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 06 03:34:49 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	237566	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	480981	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	443543	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	133480	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	204398	53.064	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	106.120%
35) Dibromofluoromethane	8.171	113	180662	52.267	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	104.540%
50) Toluene-d8	10.571	98	559732	44.755	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	89.500%
62) 4-Bromofluorobenzene	12.853	95	176568	43.992	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	87.980%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100923\
Data File : VN079476.D
Acq On : 09 Oct 2023 16:48
Operator : JC\MD
Sample : 04740-03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW3

Quant Time: Oct 09 23:17:35 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100523W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 06 03:34:49 2023
Response via : Initial Calibration

