

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100923\
 Data File : VN079478.D
 Acq On : 09 Oct 2023 17:36
 Operator : JC\MD
 Sample : 04740-05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW4

Quant Time: Oct 09 23:18:23 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100523W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 06 03:34:49 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	242185	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	478908	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	446352	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	137692	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	209517	53.355	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 106.720%		
35) Dibromofluoromethane	8.171	113	182595	53.055	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 106.120%		
50) Toluene-d8	10.571	98	570846	45.841	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 91.680%		
62) 4-Bromofluorobenzene	12.853	95	185202	46.343	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 92.680%		
Target Compounds						
						Qvalue
27) cis-1,2-Dichloroethene	7.500	96	4357	1.200	ug/l	98
44) Trichloroethene	9.359	130	7426	2.057	ug/l	89
65) Chlorobenzene	11.900	112	31521	3.171	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100923\
Data File : VN079478.D
Acq On : 09 Oct 2023 17:36
Operator : JC\MD
Sample : 04740-05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW4

Quant Time: Oct 09 23:18:23 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100523W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 06 03:34:49 2023
Response via : Initial Calibration

