

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100924\
 Data File : VN084363.D
 Acq On : 09 Oct 2024 13:33
 Operator : JC\MD
 Sample : P4340-08
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 1B-3

Manual Integrations
 APPROVED

Reviewed By : John Carlone 10/10/2024
 Supervised By : Mahesh Dadoda 10/10/2024

Quant Time: Oct 10 01:19:27 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 01 07:11:01 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	182585	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	320834	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	289609	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	110676	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	133610	49.354	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	98.700%		
35) Dibromofluoromethane	8.165	113	103935	49.139	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	98.280%		
50) Toluene-d8	10.565	98	395289	50.796	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	101.600%		
62) 4-Bromofluorobenzene	12.847	95	135162	47.676	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	95.360%		
Target Compounds						
16) Acetone	4.436	43	18063	15.532	ug/l	97
20) Methylene Chloride	5.283	84	6578	2.783	ug/l #	77
43) Isopropyl Acetate	8.688	43	131570m	21.311	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100924\
Data File : VN084363.D
Acq On : 09 Oct 2024 13:33
Operator : JC\MD
Sample : P4340-08
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
1B-3

Manual Integrations
APPROVED

Reviewed By :John Carlone 10/10/2024
Supervised By :Mahesh Dadoda 10/10/2024

Quant Time: Oct 10 01:19:27 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 01 07:11:01 2024
Response via : Initial Calibration

