

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100924\
 Data File : VN084364.D
 Acq On : 09 Oct 2024 13:57
 Operator : JC\MD
 Sample : P4353-06
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 CF-400-COMP-43

Manual Integrations
 APPROVED

Reviewed By : John Carlone 10/10/2024
 Supervised By : Mahesh Dadoda 10/10/2024

Quant Time: Oct 10 01:19:53 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 01 07:11:01 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	172805	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	314594	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	283433	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	107381	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	128180	50.028	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 100.060%			
35) Dibromofluoromethane	8.165	113	99812	48.125	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 96.260%			
50) Toluene-d8	10.565	98	384886	50.441	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 100.880%			
62) 4-Bromofluorobenzene	12.847	95	129838	46.707	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 93.420%			
Target Compounds						
16) Acetone	4.430	43	16510	15.000	ug/l	95
20) Methylene Chloride	5.271	84	6147	2.747	ug/l #	83
25) 2-Butanone	7.489	43	7591	5.064	ug/l	90
43) Isopropyl Acetate	8.688	43	151819m	25.079	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100924\
Data File : VN084364.D
Acq On : 09 Oct 2024 13:57
Operator : JC\MD
Sample : P4353-06
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
CF-400-COMP-43

Manual Integrations
APPROVED

Reviewed By : John Carlone 10/10/2024
Supervised By : Mahesh Dadoda 10/10/2024

Quant Time: Oct 10 01:19:53 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 01 07:11:01 2024
Response via : Initial Calibration

