



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100925\
Data File : VN088047.D
Acq On : 09 Oct 2025 08:42
Operator : JC\MD
Sample : BFB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT3.P

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N100825W.M
Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
Last Update : Thu Oct 09 02:36:32 2025



AutoFind: Averaged scan 1841 to 1866; Bkg corrected with scan 1840

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	20.5	3630	PASS
75	95	30	60	53.5	9489	PASS
95	95	100	100	100.0	17735	PASS
96	95	5	9	7.4	1321	PASS
173	174	0.00	2	1.1	139	PASS
174	95	50	100	68.8	12205	PASS
175	174	5	9	8.1	988	PASS
176	174	95	101	95.9	11704	PASS
177	176	5	9	6.8	799	PASS