

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN101018\
 Data File : VN051787.D
 Acq On : 10 Oct 2018 13:41
 Operator : MD\SY
 Sample : J5165-18 10X
 Misc : 5.51g/5mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 12 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 EP1-MW-A(8-12)

Quant Time: Oct 11 02:33:53 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N100418W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 09 06:03:49 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	775244	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1112721	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1057153	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	651059	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	449006	43.84	ug/l	0.00
Spiked Amount	50.000		Recovery	=	87.68%	
35) Dibromofluoromethane	7.59	113	412108	47.90	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.80%	
50) Toluene-d8	10.09	98	1602582	49.93	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.86%	
62) 4-Bromofluorobenzene	12.40	95	691729	60.89	ug/l	0.00
Spiked Amount	50.000		Recovery	=	121.78%	
Target Compounds						
39) Methylcyclohexane	9.08	83	68905	4.720	ug/l	98
73) Isopropylbenzene	12.25	105	2312221	41.663	ug/l	99
78) n-propylbenzene	12.59	91	6474225	106.953	ug/l	100
83) tert-Butylbenzene	12.99	119	453230	10.815	ug/l	98
85) sec-Butylbenzene	13.17	105	3591107	64.779	ug/l #	74
89) n-Butylbenzene	13.62	91	2894983	74.503	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN101018\
Data File : VN051787.D
Acq On : 10 Oct 2018 13:41
Operator : MD\SY
Sample : J5165-18 10X
Misc : 5.51µ/5mL/100µL/5.00mL/MSVOA_N/MEOH
ALS Vial : 12 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
EP1-MW-A(8-12)

Quant Time: Oct 11 02:33:53 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N100418W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 09 06:03:49 2018
Response via : Initial Calibration

