

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN101018\
 Data File : VN051792.D
 Acq On : 10 Oct 2018 15:45
 Operator : MD\SY
 Sample : PB113684ZHE#03
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 17 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 PB113684ZHE#03

Manual Integrations
 APPROVED

MMDadoda
 10/11/2018 5:51:19 PM

Quant Time: Oct 11 03:43:16 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N100418W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 09 06:03:49 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	813776	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1168227	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1047999	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	413447	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	465407	43.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.58%	
35) Dibromofluoromethane	7.59	113	432408	47.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.76%	
50) Toluene-d8	10.09	98	1655622	49.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.28%	
62) 4-Bromofluorobenzene	12.40	95	484447	40.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	81.24%	
Target Compounds						
18) Methyl Acetate	4.33	43	8268	1.393	ug/l #	77
20) Methylene Chloride	4.56	84	9301m	1.052	ug/l	
43) Isopropyl Acetate	8.17	43	298866	22.508	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN101018\
Data File : VN051792.D
Acq On : 10 Oct 2018 15:45
Operator : MD\SY
Sample : PB113684ZHE#03
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 17 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
PB113684ZHE#03

Manual Integrations
APPROVED

MMDadoda
10/11/2018 5:51:19 PM

Quant Time: Oct 11 03:43:16 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N100418W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 09 06:03:49 2018
Response via : Initial Calibration

