

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN101018\
 Data File : VN051798.D
 Acq On : 10 Oct 2018 18:13
 Operator : MD\SY
 Sample : PB113684ZHE#09
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 23 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 PB113684ZHE#09

Quant Time: Oct 11 04:04:17 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N100418W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 09 06:03:49 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	786964	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1143916	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1010648	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	374787	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	460510	44.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.60%	
35) Dibromofluoromethane	7.59	113	419998	47.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.98%	
50) Toluene-d8	10.09	98	1616026	48.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.96%	
62) 4-Bromofluorobenzene	12.40	95	464455	39.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	79.54%	
Target Compounds						
18) Methyl Acetate	4.34	43	12730	2.218	ug/l #	72
43) Isopropyl Acetate	8.17	43	389663	29.970	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN101018\
Data File : VN051798.D
Acq On : 10 Oct 2018 18:13
Operator : MD\SY
Sample : PB113684ZHE#09
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 23 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
PB113684ZHE#09

Quant Time: Oct 11 04:04:17 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N100418W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 09 06:03:49 2018
Response via : Initial Calibration

