

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN101018\
 Data File : VN051800.D
 Acq On : 10 Oct 2018 19:03
 Operator : MD\SY
 Sample : PB113684ZHE#11
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 25 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 PB113684ZHE#11

Manual Integrations
 APPROVED

MMDadoda
 10/11/2018 5:51:24 PM

Quant Time: Oct 11 04:09:55 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N100418W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 09 06:03:49 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	778440	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1121594	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1030732	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	410286	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	456920	44.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.86%	
35) Dibromofluoromethane	7.59	113	416181	47.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.98%	
50) Toluene-d8	10.09	98	1602254	49.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.06%	
62) 4-Bromofluorobenzene	12.40	95	475026	41.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	82.96%	
Target Compounds						
18) Methyl Acetate	4.34	43	16953m	2.987	ug/l	
20) Methylene Chloride	4.55	84	8832	1.044	ug/l #	85
43) Isopropyl Acetate	8.17	43	396005	31.064	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN101018\
Data File : VN051800.D
Acq On : 10 Oct 2018 19:03
Operator : MD\SY
Sample : PB113684ZHE#11
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 25 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
PB113684ZHE#11

Manual Integrations
APPROVED

MMDadoda
10/11/2018 5:51:24 PM

Quant Time: Oct 11 04:09:55 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N100418W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 09 06:03:49 2018
Response via : Initial Calibration

