

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN101019\
 Data File : VN058653.D
 Acq On : 9 Oct 2019 20:16
 Operator : JC/SP
 Sample : K5238-06
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-COMP

Quant Time: Oct 10 05:20:37 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N092719W.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 30 08:07:38 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	448242	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	713669	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	639119	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	262623	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	316676	48.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.70%	
35) Dibromofluoromethane	7.57	113	233776	53.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.90%	
50) Toluene-d8	10.08	98	919574	53.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.34%	
62) 4-Bromofluorobenzene	12.40	95	302548	48.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.66%	
Target Compounds						
20) Methylene Chloride	4.53	84	3440741	667.711	ug/l	97
43) Isopropyl Acetate	8.15	43	106114	9.753	ug/l	99
95) Naphthalene	15.13	128	38343	2.961	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN101019\
Data File : VN058653.D
Acq On : 9 Oct 2019 20:16
Operator : JC/SP
Sample : K5238-06
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-COMP

Quant Time: Oct 10 05:20:37 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N092719W.M
Quant Title : SW846 8260
QLast Update : Mon Sep 30 08:07:38 2019
Response via : Initial Calibration

