

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN101019\
 Data File : VN058666.D
 Acq On : 10 Oct 2019 1:03
 Operator : JC/SP
 Sample : K5261-01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 S49-0555

Quant Time: Oct 10 05:58:25 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N092719W.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 30 08:07:38 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	444097	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	694980	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	586877	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	214701	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	302669	46.64	ug/l	0.00
Spiked Amount	50.000		Recovery	=	93.28%	
35) Dibromofluoromethane	7.57	113	214985	50.47	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.94%	
50) Toluene-d8	10.09	98	845086	50.65	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.30%	
62) 4-Bromofluorobenzene	12.40	95	264080	43.77	ug/l	0.00
Spiked Amount	50.000		Recovery	=	87.54%	
Target Compounds						
3) Chloromethane	2.04	50	2864	0.535	ug/l	92
4) Vinyl Chloride	2.17	62	41485	7.911	ug/l	97
12) 1,1-Dichloroethene	3.72	96	2327	0.555	ug/l #	73
21) trans-1,2-Dichloroethene	5.01	96	96732	20.764	ug/l	97
27) cis-1,2-Dichloroethene	6.81	96	243208	44.667	ug/l	88
44) Trichloroethene	8.82	130	71241	15.104	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN101019\
Data File : VN058666.D
Acq On : 10 Oct 2019 1:03
Operator : JC/SP
Sample : K5261-01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
S49-0555

Quant Time: Oct 10 05:58:25 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N092719W.M
Quant Title : SW846 8260
QLast Update : Mon Sep 30 08:07:38 2019
Response via : Initial Calibration

