

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN101019\
 Data File : VN058644.D
 Acq On : 9 Oct 2019 16:58
 Operator : JC/SP
 Sample : K5184-10
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TB-100219

Quant Time: Oct 10 04:54:51 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N092719W.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 30 08:07:38 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	416621	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	655816	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	576693	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	229063	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	295487	48.53	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.06%	
35) Dibromofluoromethane	7.57	113	212838	52.95	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.90%	
50) Toluene-d8	10.09	98	823671	52.31	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.62%	
62) 4-Bromofluorobenzene	12.41	95	272237	47.82	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.64%	


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN101019\
Data File : VN058644.D
Acq On : 9 Oct 2019 16:58
Operator : JC/SP
Sample : K5184-10
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TB-100219

Quant Time: Oct 10 04:54:51 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N092719W.M
Quant Title : SW846 8260
QLast Update : Mon Sep 30 08:07:38 2019
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.65 min Scan# 1828

Delta R.T. 0.00 min

Lab File: VN058644.D

Acq: 9 Oct 2019 16:58

Instrument :

MSVOA_N

ClientSampled :


TB-100219

Tot Ion:168 Resp: 416621

Ion Ratio Lower Upper

168 100

99 60.0 47.7 71.5



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VNO

7.65

200000



150000

100000

50000

0

Time-->



#33

1,2-Dichloroethane-d4

Concen: 48.534 ug/l

RT: 8.01 min Scan# 1940

Delta R.T. 0.00 min

Lab File: VN058644.D


Acq: 9 Oct 2019 16:58

Tgt Ion: 65 Resp: 295487

Ion Ratio Lower Upper

65 100

67 50.6 0.0 102.4



Abundance Ion 64.90 (64.60 to 65.60): VNO

Ion 66.90 (66.60 to 67.60): VNO

8.01


150000


100000

50000

0

Time-->









#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 8.57 min Scan# 2115
 Delta R.T. 0.00 min
 Lab File: VN058644.D
 Acq: 9 Oct 2019 16:58




Instrument :
 MSVOA_N
 ClientSampleId :
 TB-100219


Tot Ion:114 Resp: 655816
 Ion Ratio Lower Upper
 114 100
 63 22.3 0.0 45.0
 88 16.2 0.0 32.2



#35
 Dibromofluoromethane
 Concen: 52.952 ug/l
 RT: 7.57 min Scan# 1804
 Delta R.T. 0.00 min
 Lab File: VN058644.D
 Acq: 9 Oct 2019 16:58

Tgt Ion:113 Resp: 212838
 Ion Ratio Lower Upper
 113 100
 111 102.6 83.4 125.2
 192 18.8 14.9 22.3





#50

Toluene-d8

Concen: 52.311 ug/l

RT: 10.09 min Scan# 2585

Delta R.T. 0.00 min

Lab File: VN058644.D

Acq: 9 Oct 2019 16:58

Instrument :

MSVOA_N

ClientSampled :





TB-100219

Tot Ion: 98 Resp: 823671

Ion Ratio Lower Upper

98 100

100 63.7 51.0 76.6



#62

4-Bromofluorobenzene

Concen: 47.816 ug/l

RT: 12.41 min Scan# 3307

Delta R.T. 0.00 min

Lab File: VN058644.D

Acq: 9 Oct 2019 16:58




Tgt Ion: 95 Resp: 272237


Ion Ratio Lower Upper

95 100

174 75.4 0.0 150.6

176 73.5 0.0 146.2





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.41 min Scan# 2996

Delta R.T. 0.00 min

Lab File: VN058644.D

Acq: 9 Oct 2019 16:58

Instrument :

MSVOA_N

ClientSampleId :

TB-100219


Tot Ion:117 Resp: 576693

Ion Ratio Lower Upper

117 100

82 57.6 47.4 71.0




119 31.0 25.5 38.3



Abundance Ion 116.90 (116.60 to 117.60): V

Ion 82.00 (81.70 to 82.70): VN

Ion 118.90 (118.60 to 119.60): V



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.35 min Scan# 3600

Delta R.T. 0.00 min

Lab File: VN058644.D

Acq: 9 Oct 2019 16:58


Tgt Ion:152 Resp: 229063

Ion Ratio Lower Upper

152 100

115 60.3 30.0 90.0

150 157.4 0.0 348.2



Abundance Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V

