

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101022\
 Data File : VN074833.D
 Acq On : 10 Oct 2022 15:48
 Operator : JC\MD
 Sample : N4894-05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-2

Quant Time: Oct 11 06:25:24 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100522W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 06 05:37:50 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	222214	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.112	114	440187	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	449807	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	210949	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	324210	61.576	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 123.160%	
35) Dibromofluoromethane	8.177	113	256954	57.775	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 115.560%	
50) Toluene-d8	10.571	98	1006534	52.744	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 105.480%	
62) 4-Bromofluorobenzene	12.853	95	249318	42.978	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 85.960%	
Target Compounds						
					Qvalue	
16) Acetone	4.471	43	11498	3.853	ug/l #	65
25) 2-Butanone	7.524	43	2113	0.390	ug/l #	82
65) Chlorobenzene	11.894	112	8224	0.597	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101022\
Data File : VN074833.D
Acq On : 10 Oct 2022 15:48
Operator : JC\MD
Sample : N4894-05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-2

Quant Time: Oct 11 06:25:24 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100522W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 06 05:37:50 2022
Response via : Initial Calibration

