

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101022\
 Data File : VN074834.D
 Acq On : 10 Oct 2022 16:12
 Operator : JC\MD
 Sample : N4894-07
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-4

Quant Time: Oct 11 06:25:51 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100522W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 06 05:37:50 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.236	168	231554	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	439336	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	446798	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	212928	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	332056	60.523	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 121.040%	
35) Dibromofluoromethane	8.177	113	257079	57.916	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 115.840%	
50) Toluene-d8	10.571	98	1028643	54.006	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 108.020%	
62) 4-Bromofluorobenzene	12.853	95	255544	44.136	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 88.280%	
Target Compounds						
					Qvalue	
65) Chlorobenzene	11.894	112	7287	0.532	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101022\
Data File : VN074834.D
Acq On : 10 Oct 2022 16:12
Operator : JC\MD
Sample : N4894-07
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-4

Quant Time: Oct 11 06:25:51 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100522W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 06 05:37:50 2022
Response via : Initial Calibration

