

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101022\
 Data File : VN074835.D
 Acq On : 10 Oct 2022 16:36
 Operator : JC\MD
 Sample : N5042-02 50X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 EOR-0614

Quant Time: Oct 11 06:26:18 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100522W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 06 05:37:50 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.236	168	378325	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	9.106	114	665006	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	669030	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	312941	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	436657	48.712	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	97.420%
35) Dibromofluoromethane	8.177	113	343856	51.177	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.360%
50) Toluene-d8	10.571	98	1469337	50.965	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.940%
62) 4-Bromofluorobenzene	12.853	95	374284	42.707	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	85.420%
Target Compounds						
16) Acetone	4.453	43	561192	110.472	ug/l	Qvalue 90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101022\
Data File : VN074835.D
Acq On : 10 Oct 2022 16:36
Operator : JC\MD
Sample : N5042-02 50X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
EOR-0614

Quant Time: Oct 11 06:26:18 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100522W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 06 05:37:50 2022
Response via : Initial Calibration

