

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101022\
 Data File : VN074836.D
 Acq On : 10 Oct 2022 17:01
 Operator : JC\MD
 Sample : N5042-03 50X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 EOR-0615

Quant Time: Oct 11 06:26:44 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100522W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 06 05:37:50 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	397406	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.112	114	684389	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	704155	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	322219	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	462111	49.076	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	98.160%
35) Dibromofluoromethane	8.177	113	364076	52.652	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	105.300%
50) Toluene-d8	10.571	98	1493507	50.336	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.680%
62) 4-Bromofluorobenzene	12.853	95	386985	42.906	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	85.820%
Target Compounds						
					Qvalue	
13) Acrolein	4.230	56	18069	3.844	ug/l	86
16) Acetone	4.453	43	10111391	1894.876	ug/l	98
25) 2-Butanone	7.500	43	36816	3.802	ug/l #	86
52) Toluene	10.635	92	9432	0.433	ug/l #	71

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101022\
Data File : VN074836.D
Acq On : 10 Oct 2022 17:01
Operator : JC\MD
Sample : N5042-03 50X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
EOR-0615

Quant Time: Oct 11 06:26:44 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100522W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 06 05:37:50 2022
Response via : Initial Calibration

