

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101022\
 Data File : VN074838.D
 Acq On : 10 Oct 2022 17:49
 Operator : JC\MD
 Sample : PB148097ZHE#06
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB148097ZHE#06

Quant Time: Oct 11 06:27:40 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100522W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 06 05:37:50 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.236	168	361327	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	637872	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	695495	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	319333	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.589	65	410437	47.941	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	95.880%		
35) Dibromofluoromethane	8.177	113	333372	51.727	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	103.460%		
50) Toluene-d8	10.571	98	1415970	51.203	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	102.400%		
62) 4-Bromofluorobenzene	12.853	95	357723	42.554	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	85.100%		
Target Compounds						
16) Acetone	4.453	43	85181	17.557	ug/l	94
20) Methylene Chloride	5.295	84	38088	2.840	ug/l #	90
43) Isopropyl Acetate	8.694	43	461858	17.862	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101022\
Data File : VN074838.D
Acq On : 10 Oct 2022 17:49
Operator : JC\MD
Sample : PB148097ZHE#06
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB148097ZHE#06

Quant Time: Oct 11 06:27:40 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100522W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 06 05:37:50 2022
Response via : Initial Calibration

