

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101022\
 Data File : VN074839.D
 Acq On : 10 Oct 2022 18:13
 Operator : JC\MD
 Sample : PB148097ZHE#07
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB148097ZHE#07

Quant Time: Oct 11 06:28:07 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100522W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 06 05:37:50 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	359516	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	652648	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	701882	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	336923	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	423253	49.687	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	99.380%
35) Dibromofluoromethane	8.177	113	333948	50.644	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.280%
50) Toluene-d8	10.571	98	1470409	51.968	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	103.940%
62) 4-Bromofluorobenzene	12.853	95	376058	43.722	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	87.440%
Target Compounds						
					Qvalue	
16) Acetone	4.459	43	107433	22.255	ug/l	96
20) Methylene Chloride	5.294	84	25817	1.194	ug/l #	93
43) Isopropyl Acetate	8.700	43	467894	17.686	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101022\
Data File : VN074839.D
Acq On : 10 Oct 2022 18:13
Operator : JC\MD
Sample : PB148097ZHE#07
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB148097ZHE#07

Quant Time: Oct 11 06:28:07 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100522W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 06 05:37:50 2022
Response via : Initial Calibration

