

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101022\
 Data File : VN074841.D
 Acq On : 11 Oct 2022 07:18
 Operator : JC\MD
 Sample : PB148097ZHE#09
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB148097ZHE#09

Manual Integrations
 APPROVED

Reviewed By :Krupa Patel 10/11/2022
 Supervised By :Mahesh Dadoda 10/11/2022

Quant Time: Oct 11 14:43:16 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100522W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 06 05:37:50 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	194028	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	355307	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	334758	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	128977	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	303484	66.013	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 132.020%#			
35) Dibromofluoromethane	8.171	113	217584	60.611	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 121.220%			
50) Toluene-d8	10.571	98	751936	48.815	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 97.640%			
62) 4-Bromofluorobenzene	12.847	95	163205	34.854	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 69.700%#			
Target Compounds					Qvalue	
16) Acetone	4.500	43	96208	36.928	ug/l #	87
20) Methylene Chloride	5.300	84	14116	1.240	ug/l #	72
43) Isopropyl Acetate	8.700	43	2908079m	201.909	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101022\
Data File : VN074841.D
Acq On : 11 Oct 2022 07:18
Operator : JC\MD
Sample : PB148097ZHE#09
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB148097ZHE#09

Manual Integrations
APPROVED

Reviewed By :Krupa Patel 10/11/2022
Supervised By :Mahesh Dadoda 10/11/2022

Quant Time: Oct 11 14:43:16 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100522W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 06 05:37:50 2022
Response via : Initial Calibration

