

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101022\
 Data File : VN074842.D
 Acq On : 11 Oct 2022 07:42
 Operator : JC\MD
 Sample : PB148097ZHE#10
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB148097ZHE#10

Quant Time: Oct 11 14:49:14 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100522W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 06 05:37:50 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.236	168	364104	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	646607	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	673497	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	308037	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	420712	48.766	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	97.540%
35) Dibromofluoromethane	8.177	113	334141	51.146	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.300%
50) Toluene-d8	10.571	98	1423842	50.792	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.580%
62) 4-Bromofluorobenzene	12.853	95	363111	42.611	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	85.220%
Target Compounds						
					Qvalue	
16) Acetone	4.453	43	105004	21.478	ug/l	100
20) Methylene Chloride	5.300	84	29972	1.709	ug/l	92
43) Isopropyl Acetate	8.694	43	676197	25.798	ug/l #	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101022\
Data File : VN074842.D
Acq On : 11 Oct 2022 07:42
Operator : JC\MD
Sample : PB148097ZHE#10
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB148097ZHE#10

Quant Time: Oct 11 14:49:14 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100522W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 06 05:37:50 2022
Response via : Initial Calibration

