

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101023\
 Data File : VN079484.D
 Acq On : 10 Oct 2023 11:22
 Operator : JC\MD
 Sample : VN1010WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1010WBL01

Quant Time: Oct 11 06:14:48 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100523W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 06 03:34:49 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	269089	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	525885	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	509095	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	159287	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	222885	51.084	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	102.160%
35) Dibromofluoromethane	8.171	113	196505	51.996	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	104.000%
50) Toluene-d8	10.571	98	629978	46.070	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	92.140%
62) 4-Bromofluorobenzene	12.853	95	206510	47.059	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.120%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101023\
Data File : VN079484.D
Acq On : 10 Oct 2023 11:22
Operator : JC\MD
Sample : VN1010WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1010WBL01

Quant Time: Oct 11 06:14:48 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100523W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 06 03:34:49 2023
Response via : Initial Calibration

