

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101023\
 Data File : VN079487.D
 Acq On : 10 Oct 2023 12:48
 Operator : JC\MD
 Sample : 04740-04DL 10X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 FW9DDL

Quant Time: Oct 11 06:16:36 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100523W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 06 03:34:49 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	263210	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	504480	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	474297	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	146703	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	212846	49.873	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	99.740%		
35) Dibromofluoromethane	8.171	113	187668	51.765	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	103.540%		
50) Toluene-d8	10.571	98	594830	45.346	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	90.700%		
62) 4-Bromofluorobenzene	12.853	95	194304	46.156	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	92.320%		
Target Compounds						
27) cis-1,2-Dichloroethene	7.494	96	157738	39.975	ug/l	96
44) Trichloroethene	9.359	130	239340	62.929	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101023\
Data File : VN079487.D
Acq On : 10 Oct 2023 12:48
Operator : JC\MD
Sample : 04740-04DL 10X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
FW9DDL

Quant Time: Oct 11 06:16:36 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100523W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 06 03:34:49 2023
Response via : Initial Calibration

