

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101023\
 Data File : VN079499.D
 Acq On : 10 Oct 2023 17:37
 Operator : JC\MD
 Sample : 04817-03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 72-11930

Quant Time: Oct 11 06:21:42 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100523W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 06 03:34:49 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	316662	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	607861	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	584388	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	226176	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	240593	46.859	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	93.720%		
35) Dibromofluoromethane	8.171	113	212904	48.738	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	97.480%		
50) Toluene-d8	10.571	98	716851	45.354	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	90.700%		
62) 4-Bromofluorobenzene	12.853	95	269079	53.048	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	106.100%		
Target Compounds					Qvalue	
16) Acetone	4.436	43	51427	25.253	ug/l #	79
20) Methylene Chloride	5.283	84	23505	4.725	ug/l #	88
43) Isopropyl Acetate	8.694	43	333953	29.219	ug/l #	71
95) Naphthalene	15.641	128	124463	16.487	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101023\
Data File : VN079499.D
Acq On : 10 Oct 2023 17:37
Operator : JC\MD
Sample : 04817-03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
72-11930

Quant Time: Oct 11 06:21:42 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100523W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 06 03:34:49 2023
Response via : Initial Calibration

