

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101023\
 Data File : VN079501.D
 Acq On : 10 Oct 2023 18:25
 Operator : JC\MD
 Sample : PB156250TB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB156250TB

Quant Time: Oct 11 06:22:35 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100523W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 06 03:34:49 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	310841	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	614276	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	595770	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	206915	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	239543	47.528	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	95.060%
35) Dibromofluoromethane	8.171	113	214402	48.569	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.140%
50) Toluene-d8	10.571	98	733530	45.924	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	91.840%
62) 4-Bromofluorobenzene	12.853	95	264638	51.627	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	103.260%
Target Compounds						
					Qvalue	
16) Acetone	4.448	43	27432	13.723	ug/l	93
20) Methylene Chloride	5.289	84	22000	4.477	ug/l #	83
43) Isopropyl Acetate	8.694	43	387434	33.545	ug/l #	70
95) Naphthalene	15.641	128	9015	1.305	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101023\
Data File : VN079501.D
Acq On : 10 Oct 2023 18:25
Operator : JC\MD
Sample : PB156250TB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB156250TB

Quant Time: Oct 11 06:22:35 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100523W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 06 03:34:49 2023
Response via : Initial Calibration

