

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101023\
 Data File : VN079506.D
 Acq On : 10 Oct 2023 20:25
 Operator : JC\MD
 Sample : 04813-06
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SB-20

Quant Time: Oct 11 06:24:43 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100523W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 06 03:34:49 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	315669	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	621696	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	606421	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	245409	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	241238	47.132	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	94.260%		
35) Dibromofluoromethane	8.171	113	214758	48.069	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	96.140%		
50) Toluene-d8	10.571	98	739097	45.720	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	91.440%		
62) 4-Bromofluorobenzene	12.853	95	276451	53.288	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	106.580%		
Target Compounds						
16) Acetone	4.436	43	69847	34.406	ug/l	90
20) Methylene Chloride	5.271	84	21518	4.290	ug/l #	87
40) Benzene	8.612	78	1409327	70.771	ug/l	100
43) Isopropyl Acetate	8.694	43	308773	26.415	ug/l #	73
52) Toluene	10.635	92	785396	67.001	ug/l	98
67) Ethyl Benzene	11.965	91	102313	4.359	ug/l	96
68) m/p-Xylenes	12.071	106	201059	23.574	ug/l	96
69) o-Xylene	12.400	106	131651	15.696	ug/l	92
80) 1,3,5-Trimethylbenzene	13.176	105	25074	1.393	ug/l	94
84) 1,2,4-Trimethylbenzene	13.488	105	72177	4.201	ug/l	97
95) Naphthalene	15.647	128	435478	53.164	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101023\
Data File : VN079506.D
Acq On : 10 Oct 2023 20:25
Operator : JC\MD
Sample : 04813-06
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SB-20

Quant Time: Oct 11 06:24:43 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100523W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 06 03:34:49 2023
Response via : Initial Calibration

