

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN101118\
 Data File : VN051815.D
 Acq On : 11 Oct 2018 16:40
 Operator : MD\SY
 Sample : J5326-45
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 FL-PH-3

Quant Time: Oct 11 17:55:03 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N101118W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 11 13:17:04 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1049950	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1463615	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1208643	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	380192	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	483550	48.18	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.36%	
35) Dibromofluoromethane	7.59	113	458043	50.33	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.66%	
50) Toluene-d8	10.09	98	1763342	52.00	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.00%	
62) 4-Bromofluorobenzene	12.40	95	463039	39.55	ug/l	0.00
Spiked Amount	50.000		Recovery	=	79.10%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN101118\
Data File : VN051815.D
Acq On : 11 Oct 2018 16:40
Operator : MD\SY
Sample : J5326-45
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
FL-PH-3

Quant Time: Oct 11 17:55:03 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N101118W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 11 13:17:04 2018
Response via : Initial Calibration

