

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN101118\
 Data File : VN051820.D
 Acq On : 11 Oct 2018 18:44
 Operator : MD\SY
 Sample : J5375-04
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 0691-KM-PA-AMD-COMP

Quant Time: Oct 12 07:00:28 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N101118W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 11 13:17:04 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1090959	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1519199	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1255663	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	412795	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	503019	48.24	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.48%	
35) Dibromofluoromethane	7.59	113	434417	45.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.98%	
50) Toluene-d8	10.09	98	1814086	51.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.08%	
62) 4-Bromofluorobenzene	12.40	95	483328	39.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	79.54%	
Target Compounds						
16) Acetone	3.82	43	22911	10.19	ug/l	98
20) Methylene Chloride	4.55	84	12314	1.32	ug/l #	93
95) Naphthalene	15.13	128	12953	5.17	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN101118\
Data File : VN051820.D
Acq On : 11 Oct 2018 18:44
Operator : MD\SY
Sample : J5375-04
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
0691-KM-PA-AMD-COMP

Quant Time: Oct 12 07:00:28 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N101118W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 11 13:17:04 2018
Response via : Initial Calibration

