

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN101118\
 Data File : VN051821.D
 Acq On : 11 Oct 2018 19:09
 Operator : MD\SY
 Sample : J5455-05
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MS-OILY-WATER

Quant Time: Oct 15 13:55:21 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N101118W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 12 09:51:20 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1076210	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1516206	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1296243	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	460012	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	508707	49.45	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.90%	
35) Dibromofluoromethane	7.59	113	469235	49.77	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.54%	
50) Toluene-d8	10.09	98	1800395	51.25	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.50%	
62) 4-Bromofluorobenzene	12.40	95	522845	43.11	ug/l	0.00
Spiked Amount	50.000		Recovery	=	86.22%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN101118\
Data File : VN051821.D
Acq On : 11 Oct 2018 19:09
Operator : MD\SY
Sample : J5455-05
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MS-OILY-WATER

Quant Time: Oct 15 13:55:21 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N101118W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 12 09:51:20 2018
Response via : Initial Calibration

