

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN101119\
 Data File : VN058685.D
 Acq On : 10 Oct 2019 23:20
 Operator : JC/SP
 Sample : VSTDCCC020
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC020

Quant Time: Oct 11 09:08:40 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N101119W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Oct 11 04:26:47 2019
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Bromochloromethane	1.000	1.000	0.0	98	0.00
2 M	Dichlorodifluoromethane	1.625	1.777	-9.4	102	0.00
3 M	Chloromethane	2.171	2.341	-7.8	102	0.00
4 M	Vinyl Chloride	2.244	2.427	-8.2	101	0.00
5 M	Bromomethane	1.386	1.433	-3.4	95	0.00
6 M	Chloroethane	1.411	1.467	-4.0	98	0.00
7 M	Trichlorofluoromethane	2.762	2.946	-6.7	101	0.00
8 T	Diethyl Ether	1.053	1.105	-4.9	99	0.00
9	1,1,2-Trichlorotrifluoroeth	1.598	1.665	-4.2	99	0.00
10 M	1,1-Dichloroethene	1.570	1.627	-3.6	98	0.00
11	Methyl Iodide	2.186	2.338	-7.0	104	0.00
12	Methyl Acetate	2.511	2.636	-5.0	103	0.00
13 M	Acrolein	0.277	0.262	5.4	96	0.00
14 M	Acrylonitrile	0.990	1.011	-2.1	101	0.00
15 M	Acetone	0.294	0.260	11.6	89	0.00
16 M	Carbon Disulfide	4.578	4.649	-1.6	96	0.00
17	Allyl chloride	3.124	2.880	7.8	88	0.00
18 M	Methylene Chloride	1.901	2.026	-6.6	101	0.00
19 M	trans-1,2-Dichloroethene	1.709	1.788	-4.6	98	0.00
20 T	Diisopropyl ether	6.365	6.637	-4.3	100	0.00
21 M	1,1-Dichloroethane	3.525	3.759	-6.6	102	0.00
22 M	cis-1,2-Dichloroethene	2.037	2.119	-4.0	98	0.00
23 M	tert-Butyl Alcohol	0.405	0.360	11.1	88	0.00
24 M	Methyl tert-Butyl Ether	5.856	6.190	-5.7	101	0.00
25 M	Chloroform	3.515	3.709	-5.5	100	0.00
26	Cyclohexane	3.094	3.189	-3.1	100	0.00
27 s	1,2-Dichloroethane-d4	2.546	2.578	-1.3	98	0.00
28 I	1,4-Difluorobenzene	1.000	1.000	0.0	98	0.00
29	1,1-Dichloropropene	0.472	0.489	-3.6	100	0.00
30 M	2-Butanone	0.267	0.245	8.2	91	0.00
31	2,2-Dichloropropane	0.557	0.498	10.6	89	0.00
32 M	1,1,1-Trichloroethane	0.546	0.569	-4.2	99	0.00
33 M	Carbon Tetrachloride	0.473	0.487	-3.0	98	0.00
34 M	Benzene	1.390	1.441	-3.7	100	0.00
35	Methacrylonitrile	0.236	0.181	23.3	74	0.00
36 M	1,2-Dichloroethane	0.532	0.557	-4.7	100	0.00
37 M	Trichloroethene	0.342	0.356	-4.1	100	0.00
38	Methylcyclohexane	0.548	0.547	0.2	98	0.00
39 M	1,2-Dichloropropane	0.380	0.394	-3.7	98	0.00
40	Dibromomethane	0.240	0.246	-2.5	98	0.00
41 M	Bromodichloromethane	0.490	0.502	-2.4	98	0.00
42 M	Vinyl Acetate	0.935	0.970	-3.7	96	0.00
43	Ethyl Acetate	0.524	0.509	2.9	93	0.00
44	Isopropyl Acetate	0.848	0.853	-0.6	98	0.00
45 T	1,4-Dioxane	0.007	0.005#	28.6	82	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN101119\
 Data File : VN058685.D
 Acq On : 10 Oct 2019 23:20
 Operator : JC/SP
 Sample : VSTDCCC020
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC020

Quant Time: Oct 11 09:08:40 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N101119W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Oct 11 04:26:47 2019
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
46	Methyl methacrylate	0.390	0.393	-0.8	101	0.00
47	n-amyl Acetate	0.752	0.745	0.9	99	0.00
48 M	t-1,3-Dichloropropene	0.568	0.566	0.4	97	0.00
49 T	cis-1,3-Dichloropropene	0.597	0.602	-0.8	97	0.00
50 M	1,1,2-Trichloroethane	0.342	0.356	-4.1	100	0.00
51	Ethyl methacrylate	0.548	0.544	0.7	99	0.00
52	1,3-Dichloropropane	0.610	0.632	-3.6	99	0.00
53 M	Dibromochloromethane	0.367	0.368	-0.3	98	0.00
54 M	1,2-Dibromoethane	0.355	0.365	-2.8	100	0.00
55 M	2-Chloroethyl vinyl ether	0.272	0.286	-5.1	97	0.00
56 M	Bromoform	0.256	0.248	3.1	97	0.00
57 I	Chlorobenzene-d5	1.000	1.000	0.0	99	0.00
58 M	4-Methyl-2-Pentanone	0.547	0.575	-5.1	97	0.00
59 M	2-Hexanone	0.429	0.417	2.8	93	0.00
60 S	4-Bromofluorobenzene	0.525	0.524	0.2	99	0.00
61 M	Tetrachloroethene	0.324	0.358	-10.5	104	0.00
62 M	Toluene	1.614	1.688	-4.6	99	0.00
63 S	Toluene-d8	1.416	1.441	-1.8	99	0.00
64 M	Chlorobenzene	0.987	1.019	-3.2	99	0.00
65	1,1,1,2-Tetrachloroethane	0.370	0.390	-5.4	101	0.00
66 M	Ethyl Benzene	1.787	1.906	-6.7	101	0.00
67 M	m/p-Xylenes	0.668	0.694	-3.9	100	0.00
68 M	o-Xylene	0.647	0.671	-3.7	100	0.00
69 M	Styrene	1.069	1.097	-2.6	101	0.00
70	Isopropylbenzene	1.747	1.842	-5.4	101	0.00
71 M	1,1,2,2-Tetrachloroethane	0.588	0.588	0.0	96	0.00
72	1,2,3-Trichloropropane	0.533	0.548	-2.8	98	0.00
73	Bromobenzene	0.424	0.440	-3.8	101	0.00
74	n-propylbenzene	2.045	2.144	-4.8	99	0.00
75	2-Chlorotoluene	1.231	1.278	-3.8	100	0.00
76	1,3,5-Trimethylbenzene	1.497	1.553	-3.7	100	0.00
77	t-1,4-Dichloro-2-butene	0.203	0.178	12.3	88	0.00
78	4-Chlorotoluene	1.268	1.321	-4.2	101	0.00
79	tert-butylbenzene	1.279	1.320	-3.2	99	0.00
80	1,2,4-Trimethylbenzene	1.487	1.570	-5.6	101	0.00
81	sec-Butylbenzene	1.721	1.801	-4.6	100	0.00
82	p-Isopropyltoluene	1.545	1.604	-3.8	100	0.00
83 M	1,3-Dichlorobenzene	0.768	0.794	-3.4	101	0.00
84 M	1,4-Dichlorobenzene	0.767	0.793	-3.4	101	0.00
85	n-Butylbenzene	1.415	1.420	-0.4	100	0.00
86 T	Hexachloroethane	0.281	0.280	0.4	98	0.00
87 M	1,2-Dichlorobenzene	0.758	0.786	-3.7	102	0.00
88	1,2-Dibromo-3-Chloropropane	0.131	0.129	1.5	95	0.00
89	1,2,4-Trichlorobenzene	0.479	0.470	1.9	98	0.00
90	Hexachlorobutadiene	0.247	0.243	1.6	97	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN101119\
Data File : VN058685.D
Acq On : 10 Oct 2019 23:20
Operator : JC/SP
Sample : VSTDCCC020
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC020

Quant Time: Oct 11 09:08:40 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N101119W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Fri Oct 11 04:26:47 2019
Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
91 M	Naphthalene	1.413	1.351	4.4	95	0.00
92	1,2,3-Trichlorobenzene	0.466	0.448	3.9	95	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0