

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN101122\  
 Data File : VN074849.D  
 Acq On : 11 Oct 2022 15:54  
 Operator : JC\MD  
 Sample : VN1011WBL01  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 3 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 VN1011WBL01

Quant Time: Oct 12 03:53:15 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N100522W.M  
 Quant Title : SW846 8260  
 QLast Update : Thu Oct 06 05:37:50 2022  
 Response via : Initial Calibration


| Compound                      | R.T.           | QIon | Response   | Conc      | Units  | Dev(Min) |
|-------------------------------|----------------|------|------------|-----------|--------|----------|
| -----                         |                |      |            |           |        |          |
| Internal Standards            |                |      |            |           |        |          |
| 1) Pentafluorobenzene         | 8.235          | 168  | 406613     | 50.000    | ug/l   | 0.00     |
| 34) 1,4-Difluorobenzene       | 9.106          | 114  | 709224     | 50.000    | ug/l   | 0.00     |
| 63) Chlorobenzene-d5          | 11.870         | 117  | 712667     | 50.000    | ug/l   | 0.00     |
| 72) 1,4-Dichlorobenzene-d4    | 13.794         | 152  | 305890     | 50.000    | ug/l   | 0.00     |
|                               |                |      |            |           |        |          |
| System Monitoring Compounds   |                |      |            |           |        |          |
| 33) 1,2-Dichloroethane-d4     | 8.594          | 65   | 458634     | 47.604    | ug/l   | 0.00     |
| Spiked Amount 50.000          | Range 74 - 125 |      | Recovery = | 95.200%   |        |          |
| 35) Dibromofluoromethane      | 8.177          | 113  | 365776     | 51.045    | ug/l   | 0.00     |
| Spiked Amount 50.000          | Range 75 - 124 |      | Recovery = | 102.100%  |        |          |
| 50) Toluene-d8                | 10.570         | 98   | 1550595    | 50.430    | ug/l   | 0.00     |
| Spiked Amount 50.000          | Range 86 - 113 |      | Recovery = | 100.860%  |        |          |
| 62) 4-Bromofluorobenzene      | 12.853         | 95   | 397811     | 42.562    | ug/l   | 0.00     |
| Spiked Amount 50.000          | Range 83 - 123 |      | Recovery = | 85.120%   |        |          |
|                               |                |      |            |           |        |          |
| Target Compounds              |                |      |            |           | Qvalue |          |
| 11) Tert butyl alcohol        | 5.553          | 59   | 700        | 0.217     | ug/l # | 74       |
| 13) Acrolein                  | 4.200          | 56   | 521        | Below Cal | #      | 83       |
| 15) Acrylonitrile             | 5.741          | 53   | 1552       | 0.221     | ug/l # | 36       |
| 16) Acetone                   | 4.489          | 43   | 3892       | 0.713     | ug/l # | 44       |
| 18) Methyl Acetate            | 5.047          | 43   | 8179       | 0.405     | ug/l # | 79       |
| 20) Methylene Chloride        | 5.283          | 84   | 1411       | Below Cal | #      | 74       |
| 25) 2-Butanone                | 7.512          | 43   | 4015       | 0.405     | ug/l # | 48       |
| 31) Cyclohexane               | 8.241          | 56   | 10911      | 0.655     | ug/l # | 57       |
| 36) 1,1-Dichloropropene       | 8.235          | 75   | 36327      | 3.132     | ug/l # | 50       |
| 38) Carbon Tetrachloride      | 8.235          | 117  | 44753      | 4.092     | ug/l # | 16       |
| 49) 1,4-Dioxane               | 9.688          | 88   | 55         | 0.215     | ug/l # | 53       |
| 51) 4-Methyl-2-Pentanone      | 10.459         | 43   | 3744       | 0.202     | ug/l # | 83       |
| 58) 2-Chloroethyl Vinyl ether | 10.165         | 63   | 2551       | 0.428     | ug/l # | 68       |
| 59) 2-Hexanone                | 11.200         | 43   | 4499       | 0.321     | ug/l   | 80       |
| 71) Bromoform                 | 12.853         | 173  | 5179       | 0.852     | ug/l # | 1        |
| 74) N-amyl acetate            | 12.500         | 43   | 4977       | 0.280     | ug/l # | 78       |
| 75) 1,1,2,2-Tetrachloroethane | 12.947         | 83   | 889        | Below Cal | #      | 81       |
| 76) 1,2,3-Trichloropropane    | 12.853         | 75   | 237135     | 25.973    | ug/l # | 1        |
| 77) Bromobenzene              | 12.988         | 156  | 1042       | Below Cal |        | 94       |
| 81) trans-1,4-Dichloro-2-b... | 12.853         | 75   | 237135     | 50.605    | ug/l # | 14       |
| 87) 1,3-Dichlorobenzene       | 13.741         | 146  | 3067       | 0.263     | ug/l # | 75       |
| 88) 1,4-Dichlorobenzene       | 13.817         | 146  | 4491       | 0.407     | ug/l # | 65       |
| 91) 1,2-Dichlorobenzene       | 14.117         | 146  | 2672       | 0.238     | ug/l   | 91       |
| 93) 1,2,4-Trichlorobenzene    | 15.394         | 180  | 3875       | 0.788     | ug/l # | 51       |
| 95) Naphthalene               | 15.641         | 128  | 23732      | 1.116     | ug/l   | 96       |
| 96) 1,2,3-Trichlorobenzene    | 15.841         | 180  | 3708       | Below Cal | #      | 10       |
| -----                         |                |      |            |           |        |          |


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN101122\  
Data File : VN074849.D  
Acq On : 11 Oct 2022 15:54  
Operator : JC\MD  
Sample : VN1011WBL01  
Misc : 5.0mL/MSVOA\_N/WATER  
ALS Vial : 3 Sample Multiplier: 1

Instrument :  
MSVOA\_N  
ClientSampleId :  
VN1011WBL01

Quant Time: Oct 12 03:53:15 2022  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N100522W.M  
Quant Title : SW846 8260  
QLast Update : Thu Oct 06 05:37:50 2022  
Response via : Initial Calibration









#1  
Pentafluorobenzene  
Concen: 50.000 ug/l  
RT: 8.235 min Scan# 1051  
Delta R.T. -0.001 min  
Lab File: VN074849.D  
Acq: 11 Oct 2022 15:54




Instrument :  
MSVOA\_N  
ClientSampleId :  
VN1011WBL01


Tgt Ion:168 Resp: 406613  
Ion Ratio Lower Upper  
168 100  
99 56.1 42.8 64.2



#11  
Tert butyl alcohol  
Concen: 0.217 ug/l  
RT: 5.553 min Scan# 595  
Delta R.T. -0.006 min  
Lab File: VN074849.D  
Acq: 11 Oct 2022 15:54

Tgt Ion: 59 Resp: 700  
Ion Ratio Lower Upper  
59 100  
57 0.0 7.7 11.5#









#13  
Acrolein  
Concen: Below Cal  
RT: 4.200 min Scan# 365  
Delta R.T. -0.001 min  
Lab File: VN074849.D  
Acq: 11 Oct 2022 15:54




Instrument :  
MSVOA\_N  
ClientSampleId :  
VN1011WBL01


Tgt Ion: 56 Resp: 521  
Ion Ratio Lower Upper  
56 100  
55 82.7 54.9 82.3#



#15  
Acrylonitrile  
Concen: 0.221 ug/l  
RT: 5.741 min Scan# 627  
Delta R.T. -0.000 min  
Lab File: VN074849.D  
Acq: 11 Oct 2022 15:54

Tgt Ion: 53 Resp: 1552  
Ion Ratio Lower Upper  
53 100  
52 0.0 64.3 96.5#  
51 34.5 27.5 41.3









#16  
Acetone  
Concen: 0.713 ug/l  
RT: 4.489 min Scan# 414  
Delta R.T. 0.029 min  
Lab File: VN074849.D  
Acq: 11 Oct 2022 15:54




Instrument :  
MSVOA\_N  
ClientSampleId :  
VN1011WBL01


Tgt Ion: 43 Resp: 3892  
Ion Ratio Lower Upper  
43 100  
58 0.0 24.9 37.3#



#18  
Methyl Acetate  
Concen: 0.405 ug/l  
RT: 5.047 min Scan# 509  
Delta R.T. -0.000 min  
Lab File: VN074849.D  
Acq: 11 Oct 2022 15:54

Tgt Ion: 43 Resp: 8179  
Ion Ratio Lower Upper  
43 100  
74 13.1 18.5 27.7#









#20  
Methylene Chloride  
Concen: Below Cal  
RT: 5.283 min Scan# 549  
Delta R.T. -0.012 min  
Lab File: VN074849.D  
Acq: 11 Oct 2022 15:54




Instrument :  
MSVOA\_N  
ClientSampleId :  
VN1011WBL01


Tgt Ion: 84 Resp: 1411  
Ion Ratio Lower Upper  
84 100  
49 139.4 107.4 161.0  
51 0.0 35.1 52.7#  
86 29.9 50.4 75.6#



#25  
2-Butanone  
Concen: 0.405 ug/l  
RT: 7.512 min Scan# 928  
Delta R.T. 0.011 min  
Lab File: VN074849.D  
Acq: 11 Oct 2022 15:54

Tgt Ion: 43 Resp: 4015  
Ion Ratio Lower Upper  
43 100  
72 0.0 21.4 32.0#





#31

Cyclohexane

Concen: 0.655 ug/l

RT: 8.241 min Scan# 10911

Delta R.T. -0.024 min

Lab File: VN074849.D

Acq: 11 Oct 2022 15:54

Instrument :

MSVOA\_N

ClientSampleId :

VN1011WBL01





Tgt Ion: 56 Resp: 10911

Ion Ratio Lower Upper

56 100

69 80.5 23.1 34.7#

84 105.7 67.1 100.7#



#33

1,2-Dichloroethane-d4

Concen: 47.604 ug/l

RT: 8.594 min Scan# 1112

Delta R.T. 0.006 min

Lab File: VN074849.D




Acq: 11 Oct 2022 15:54


Tgt Ion: 65 Resp: 458634

Ion Ratio Lower Upper

65 100

67 55.2 0.0 109.0





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.106 min Scan# 1199

Delta R.T. -0.000 min

Lab File: VN074849.D

Acq: 11 Oct 2022 15:54

Instrument :

MSVOA\_N

ClientSampleId :

VN1011WBL01





Tgt Ion:114 Resp: 709224

Ion Ratio Lower Upper

114 100

63 21.2 0.0 41.2

88 14.6 0.0 30.2



#35

Dibromofluoromethane

Concen: 51.045 ug/l

RT: 8.177 min Scan# 1041

Delta R.T. 0.006 min

Lab File: VN074849.D

Acq: 11 Oct 2022 15:54




Tgt Ion:113 Resp: 365776


Ion Ratio Lower Upper

113 100

111 103.7 83.0 124.6

192 16.9 12.6 18.8









#36  
1,1-Dichloropropene  
Concen: 3.132 ug/l  
RT: 8.235 min Scan# 1051  
Delta R.T. -0.141 min  
Lab File: VN074849.D  
Acq: 11 Oct 2022 15:54




Instrument :  
MSVOA\_N  
ClientSampleId :  
VN1011WBL01


Tgt Ion: 75 Resp: 36327  
Ion Ratio Lower Upper  
75 100  
110 8.1 17.2 51.5#  
77 0.3 24.3 36.5#



#38  
Carbon Tetrachloride  
Concen: 4.092 ug/l  
RT: 8.235 min Scan# 1051  
Delta R.T. -0.136 min  
Lab File: VN074849.D  
Acq: 11 Oct 2022 15:54

Tgt Ion: 117 Resp: 44753  
Ion Ratio Lower Upper  
117 100  
119 3.5 74.3 111.5#  
121 0.0 25.0 37.4#









#49  
1,4-Dioxane  
Concen: 0.215 ug/l  
RT: 9.688 min Scan# 11  
Delta R.T. -0.018 min  
Lab File: VN074849.D  
Acq: 11 Oct 2022 15:54




Instrument :  
MSVOA\_N  
ClientSampleId :  
VN1011WBL01


Tgt Ion: 88 Resp: 55  
Ion Ratio Lower Upper  
88 100  
43 0.0 23.4 35.2#  
58 103.6 54.9 82.3#



#50  
Toluene-d8  
Concen: 50.430 ug/l  
RT: 10.570 min Scan# 1448  
Delta R.T. -0.000 min  
Lab File: VN074849.D  
Acq: 11 Oct 2022 15:54

Tgt Ion: 98 Resp: 1550595  
Ion Ratio Lower Upper  
98 100  
100 63.6 50.7 76.1









#51  
4-Methyl-2-Pentanone  
Concen: 0.202 ug/l  
RT: 10.459 min Scan# 1427  
Delta R.T. 0.011 min  
Lab File: VN074849.D  
Acq: 11 Oct 2022 15:54




Instrument :  
MSVOA\_N  
ClientSampleId :  
VN1011WBL01


Tgt Ion: 43 Resp: 3744  
Ion Ratio Lower Upper  
43 100  
58 28.1 30.6 45.8#



#58  
2-Chloroethyl Vinyl ether  
Concen: 0.428 ug/l  
RT: 10.165 min Scan# 1379  
Delta R.T. -0.000 min  
Lab File: VN074849.D  
Acq: 11 Oct 2022 15:54

Tgt Ion: 63 Resp: 2551  
Ion Ratio Lower Upper  
63 100  
106 41.1 20.0 30.0#









#59  
2-Hexanone  
Concen: 0.321 ug/l  
RT: 11.200 min Scan# 11  
Delta R.T. -0.000 min  
Lab File: VN074849.D  
Acq: 11 Oct 2022 15:54




Instrument :  
MSVOA\_N  
ClientSampleId :  
VN1011WBL01


Tgt Ion: 43 Resp: 4499  
Ion Ratio Lower Upper  
43 100  
58 69.0 27.3 81.9



#62  
4-Bromofluorobenzene  
Concen: 42.562 ug/l  
RT: 12.853 min Scan# 1836  
Delta R.T. -0.000 min  
Lab File: VN074849.D  
Acq: 11 Oct 2022 15:54

Tgt Ion: 95 Resp: 397811  
Ion Ratio Lower Upper  
95 100  
174 80.4 0.0 149.4  
176 78.1 0.0 142.0









#63  
Chlorobenzene-d5  
Concen: 50.000 ug/l  
RT: 11.870 min Scan# 11  
Delta R.T. -0.000 min  
Lab File: VN074849.D  
Acq: 11 Oct 2022 15:54




Instrument :  
MSVOA\_N  
ClientSampleId :  
VN1011WBL01


Tgt Ion:117 Resp: 712667  
Ion Ratio Lower Upper  
117 100  
82 58.2 45.4 68.2  
119 31.4 25.1 37.7



#71  
Bromoform  
Concen: 0.852 ug/l  
RT: 12.853 min Scan# 1836  
Delta R.T. 0.270 min  
Lab File: VN074849.D  
Acq: 11 Oct 2022 15:54

Tgt Ion:173 Resp: 5179  
Ion Ratio Lower Upper  
173 100  
175 395.8 23.9 71.7#  
254 0.0 0.1 0.1#









#72  
1,4-Dichlorobenzene-d4  
Concen: 50.000 ug/l  
RT: 13.794 min Scan# 1996  
Delta R.T. -0.000 min  
Lab File: VN074849.D  
Acq: 11 Oct 2022 15:54




Instrument :  
MSVOA\_N  
ClientSampleId :  
VN1011WBL01


Tgt Ion:152 Resp: 305890  
Ion Ratio Lower Upper  
152 100  
115 61.0 30.0 90.0  
150 163.1 0.0 356.4



#74  
N-ethyl acetate  
Concen: 0.280 ug/l  
RT: 12.500 min Scan# 1776  
Delta R.T. 0.006 min  
Lab File: VN074849.D  
Acq: 11 Oct 2022 15:54

Tgt Ion: 43 Resp: 4977  
Ion Ratio Lower Upper  
43 100  
70 22.0 32.9 49.3#  
55 19.3 17.8 26.6  
61 12.8 18.9 28.3#









#75  
1,1,2,2-Tetrachloroethane  
Concen: Below Cal  
RT: 12.947 min Scan# 1851  
Delta R.T. 0.006 min  
Lab File: VN074849.D  
Acq: 11 Oct 2022 15:54




Instrument :  
MSVOA\_N  
ClientSampleId :  
VN1011WBL01


Tgt Ion: 83 Resp: 889  
Ion Ratio Lower Upper  
83 100  
131 0.0 4.3 12.8#  
85 51.3 32.9 98.6



#76  
1,2,3-Trichloropropane  
Concen: 25.973 ug/l  
RT: 12.853 min Scan# 1836  
Delta R.T. -0.141 min  
Lab File: VN074849.D  
Acq: 11 Oct 2022 15:54

Tgt Ion: 75 Resp: 237135  
Ion Ratio Lower Upper  
75 100  
77 1.8 76.9 230.7#









#77  
Bromobenzene  
Concen: Below Cal  
RT: 12.988 min Scan# 1859  
Delta R.T. 0.006 min  
Lab File: VN074849.D  
Acq: 11 Oct 2022 15:54




Instrument :  
MSVOA\_N  
ClientSampleId :  
VN1011WBL01


Tgt Ion:156 Resp: 1042  
Ion Ratio Lower Upper  
156 100  
77 259.7 124.7 374.1  
158 104.2 48.3 144.9



#81  
trans-1,4-Dichloro-2-butene  
Concen: 50.605 ug/l  
RT: 12.853 min Scan# 1836  
Delta R.T. 0.117 min  
Lab File: VN074849.D  
Acq: 11 Oct 2022 15:54

Tgt Ion: 75 Resp: 237135  
Ion Ratio Lower Upper  
75 100  
53 0.2 71.4 107.0#  
89 0.1 32.1 48.1#









#87  
 1,3-Dichlorobenzene  
 Concen: 0.263 ug/l  
 RT: 13.741 min Scan# 1986  
 Delta R.T. 0.006 min  
 Lab File: VN074849.D  
 Acq: 11 Oct 2022 15:54




Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 VN1011WBL01


Tgt Ion:146 Resp: 3067  
 Ion Ratio Lower Upper  
 146 100  
 111 41.6 20.8 62.2  
 148 30.9 31.5 94.5#



#88  
 1,4-Dichlorobenzene  
 Concen: 0.407 ug/l  
 RT: 13.817 min Scan# 2000  
 Delta R.T. 0.006 min  
 Lab File: VN074849.D  
 Acq: 11 Oct 2022 15:54

Tgt Ion:146 Resp: 4491  
 Ion Ratio Lower Upper  
 146 100  
 111 0.0 19.9 59.6#  
 148 50.5 31.8 95.4





#91

1,2-Dichlorobenzene

Concen: 0.238 ug/l

RT: 14.117 min Scan# 2049

Delta R.T. 0.011 min

Lab File: VN074849.D

Acq: 11 Oct 2022 15:54

Instrument :

MSVOA\_N

ClientSampleId :

VN1011WBL01





Tgt Ion:146 Resp: 2672

Ion Ratio Lower Upper

146 100

111 32.6 21.5 64.5

148 67.7 31.9 95.9



#93

1,2,4-Trichlorobenzene

Concen: 0.788 ug/l

RT: 15.394 min Scan# 2268

Delta R.T. -0.000 min

Lab File: VN074849.D

Acq: 11 Oct 2022 15:54




Tgt Ion:180 Resp: 3875


Ion Ratio Lower Upper

180 100

182 37.0 48.5 145.5#

145 24.5 16.1 48.3





#95

Naphthalene

Concen: 1.116 ug/l

RT: 15.641 min Scan# 21

Delta R.T. -0.000 min

Lab File: VN074849.D


Acq: 11 Oct 2022 15:54

Instrument :

MSVOA\_N

ClientSampleId :

VN1011WBL01



Tgt Ion:128 Resp: 23732

Ion Ratio Lower Upper

128 100

127 14.8 10.2 15.2

129 10.4 9.0 13.4

