

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101122\
 Data File : VN074852.D
 Acq On : 11 Oct 2022 17:25
 Operator : JC\MD
 Sample : PB148255ZHE#01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB148255ZHE#01

Quant Time: Oct 12 03:55:26 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100522W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 06 05:37:50 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	374159	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.112	114	727072	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	761749	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	324590	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	271219	30.593	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	61.180%#
35) Dibromofluoromethane	8.177	113	243400	33.134	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	66.260%#
50) Toluene-d8	10.571	98	897741	28.481	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	56.960%#
62) 4-Bromofluorobenzene	12.853	95	319437	33.338	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	66.680%#
Target Compounds						
					Qvalue	
16) Acetone	4.459	43	117496	23.387	ug/l #	84
20) Methylene Chloride	5.306	84	29840	1.583	ug/l #	75
43) Isopropyl Acetate	8.700	43	499791	16.958	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101122\
Data File : VN074852.D
Acq On : 11 Oct 2022 17:25
Operator : JC\MD
Sample : PB148255ZHE#01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB148255ZHE#01

Quant Time: Oct 12 03:55:26 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100522W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 06 05:37:50 2022
Response via : Initial Calibration

