

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101122\
 Data File : VN074855.D
 Acq On : 11 Oct 2022 18:38
 Operator : JC\MD
 Sample : PB148255ZHE#04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB148255ZHE#04

Quant Time: Oct 12 03:56:57 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100522W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 06 05:37:50 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.236	168	438964	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.112	114	880330	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	932440	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	408860	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	319322	30.701	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	61.400%#
35) Dibromofluoromethane	8.177	113	281321	31.629	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	63.260%#
50) Toluene-d8	10.571	98	1058040	27.723	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	55.440%#
62) 4-Bromofluorobenzene	12.853	95	398512	34.350	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	68.700%#
Target Compounds						
43) Isopropyl Acetate	8.700	43	483615	13.552	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101122\
 Data File : VN074855.D
 Acq On : 11 Oct 2022 18:38
 Operator : JC\MD
 Sample : PB148255ZHE#04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB148255ZHE#04

Quant Time: Oct 12 03:56:57 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100522W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 06 05:37:50 2022
 Response via : Initial Calibration

