

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101122\
 Data File : VN074856.D
 Acq On : 11 Oct 2022 19:02
 Operator : JC\MD
 Sample : PB148255ZHE#05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB148255ZHE#05

Quant Time: Oct 12 03:57:25 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100522W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 06 05:37:50 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.236	168	402002	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	802341	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	850841	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	354492	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	296467	31.125	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	62.240%#
35) Dibromofluoromethane	8.177	113	253206	31.235	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	62.460%#
50) Toluene-d8	10.571	98	976590	28.076	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	56.160%#
62) 4-Bromofluorobenzene	12.853	95	354066	33.485	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	66.980%#
Target Compounds						
					Qvalue	
16) Acetone	4.465	43	104751	19.406	ug/l	97
20) Methylene Chloride	5.289	84	34550	1.887	ug/l #	93
43) Isopropyl Acetate	8.700	43	482532	14.836	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101122\
Data File : VN074856.D
Acq On : 11 Oct 2022 19:02
Operator : JC\MD
Sample : PB148255ZHE#05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB148255ZHE#05

Quant Time: Oct 12 03:57:25 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100522W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 06 05:37:50 2022
Response via : Initial Calibration

