

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101122\
 Data File : VN074857.D
 Acq On : 11 Oct 2022 19:26
 Operator : JC\MD
 Sample : PB148255ZHE#06
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB148255ZHE#06

Quant Time: Oct 12 03:57:51 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100522W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 06 05:37:50 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	418070	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.112	114	816217	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	870803	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	369123	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	291069	29.384	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	58.760%#		
35) Dibromofluoromethane	8.183	113	261967	31.766	ug/l	0.01
Spiked Amount 50.000	Range 75 - 124		Recovery =	63.540%#		
50) Toluene-d8	10.571	98	975573	27.570	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	55.140%#		
62) 4-Bromofluorobenzene	12.853	95	364124	33.851	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	67.700%#		
Target Compounds						Qvalue
43) Isopropyl Acetate	8.700	43	471832	14.261	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101122\
Data File : VN074857.D
Acq On : 11 Oct 2022 19:26
Operator : JC\MD
Sample : PB148255ZHE#06
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB148255ZHE#06

Quant Time: Oct 12 03:57:51 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100522W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 06 05:37:50 2022
Response via : Initial Calibration

