

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101122\
 Data File : VN074859.D
 Acq On : 11 Oct 2022 20:14
 Operator : JC\MD
 Sample : PB148255ZHE#08
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB148255ZHE#08

Quant Time: Oct 12 03:58:41 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100522W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 06 05:37:50 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	439526	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	863572	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	927254	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	394206	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	308469	29.620	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	59.240%#
35) Dibromofluoromethane	8.171	113	277454	31.799	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	63.600%#
50) Toluene-d8	10.571	98	1044401	27.896	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	55.800%#
62) 4-Bromofluorobenzene	12.853	95	386233	33.937	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	67.880%#
Target Compounds						
					Qvalue	
16) Acetone	4.459	43	140376	23.786	ug/l	94
43) Isopropyl Acetate	8.700	43	468933	13.396	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101122\
Data File : VN074859.D
Acq On : 11 Oct 2022 20:14
Operator : JC\MD
Sample : PB148255ZHE#08
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB148255ZHE#08

Quant Time: Oct 12 03:58:41 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100522W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 06 05:37:50 2022
Response via : Initial Calibration

