

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101122\
 Data File : VN074860.D
 Acq On : 11 Oct 2022 20:38
 Operator : JC\MD
 Sample : PB148255ZHE#09
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB148255ZHE#09

Quant Time: Oct 12 03:59:06 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100522W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 06 05:37:50 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	409988	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	810842	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	895026	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	402515	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	295449	30.414	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	60.820%#
35) Dibromofluoromethane	8.177	113	263804	32.201	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	64.400%#
50) Toluene-d8	10.570	98	982327	27.945	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	55.880%#
62) 4-Bromofluorobenzene	12.853	95	385275	36.055	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	72.100%#
Target Compounds						
					Qvalue	
16) Acetone	4.453	43	105874	19.232	ug/l	94
20) Methylene Chloride	5.300	84	33443	1.672	ug/l #	71
43) Isopropyl Acetate	8.694	43	464770	14.140	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101122\
Data File : VN074860.D
Acq On : 11 Oct 2022 20:38
Operator : JC\MD
Sample : PB148255ZHE#09
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB148255ZHE#09

Quant Time: Oct 12 03:59:06 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100522W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 06 05:37:50 2022
Response via : Initial Calibration

