

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101122\
 Data File : VN074861.D
 Acq On : 11 Oct 2022 21:03
 Operator : JC\MD
 Sample : PB148255ZHE#10
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB148255ZHE#10

Quant Time: Oct 12 03:59:32 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100522W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 06 05:37:50 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.236	168	432589	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.112	114	861622	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	934653	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	398026	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	309013	30.148	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	60.300%#
35) Dibromofluoromethane	8.177	113	283368	32.551	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	65.100%#
50) Toluene-d8	10.571	98	1029015	27.548	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	55.100%#
62) 4-Bromofluorobenzene	12.853	95	391117	34.444	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	68.880%#
Target Compounds						
					Qvalue	
16) Acetone	4.459	43	142105	24.465	ug/l	100
20) Methylene Chloride	5.295	84	36395	1.798	ug/l #	82
43) Isopropyl Acetate	8.694	43	478772	13.708	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101122\
Data File : VN074861.D
Acq On : 11 Oct 2022 21:03
Operator : JC\MD
Sample : PB148255ZHE#10
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB148255ZHE#10

Quant Time: Oct 12 03:59:32 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100522W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 06 05:37:50 2022
Response via : Initial Calibration

