

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101122\
 Data File : VN074863.D
 Acq On : 11 Oct 2022 21:51
 Operator : JC\MD
 Sample : PB148255ZHE#12
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB148255ZHE#12

Quant Time: Oct 12 04:00:28 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100522W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 06 05:37:50 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.236	168	388561	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	9.106	114	786062	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	859683	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	378941	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	286963	31.169	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	62.340%#
35) Dibromofluoromethane	8.177	113	259058	32.619	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	65.240%#
50) Toluene-d8	10.571	98	944043	27.702	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	55.400%#
62) 4-Bromofluorobenzene	12.853	95	368284	35.551	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	71.100%#
Target Compounds						
					Qvalue	
16) Acetone	4.459	43	146509	28.081	ug/l	100
20) Methylene Chloride	5.300	84	35840	2.195	ug/l	# 85
43) Isopropyl Acetate	8.700	43	462885	14.527	ug/l	# 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101122\
Data File : VN074863.D
Acq On : 11 Oct 2022 21:51
Operator : JC\MD
Sample : PB148255ZHE#12
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB148255ZHE#12

Quant Time: Oct 12 04:00:28 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100522W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 06 05:37:50 2022
Response via : Initial Calibration

