

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101122\
 Data File : VN074865.D
 Acq On : 11 Oct 2022 22:39
 Operator : JC\MD
 Sample : PB148097ZHE#15
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB148097ZHE#15

Quant Time: Oct 12 04:01:22 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100522W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 06 05:37:50 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.236	168	432652	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.112	114	859515	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	937944	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	411663	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	302303	29.489	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	58.980%#
35) Dibromofluoromethane	8.177	113	284219	32.728	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	65.460%#
50) Toluene-d8	10.571	98	1031077	27.670	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	55.340%#
62) 4-Bromofluorobenzene	12.853	95	394657	34.841	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	69.680%#
Target Compounds						
					Qvalue	
16) Acetone	4.465	43	111909	19.263	ug/l	97
20) Methylene Chloride	5.295	84	30665	1.149	ug/l #	80
43) Isopropyl Acetate	8.694	43	479292	13.756	ug/l #	92
96) 1,2,3-Trichlorobenzene	15.700	180	62	Below Cal	#	12

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101122\
Data File : VN074865.D
Acq On : 11 Oct 2022 22:39
Operator : JC\MD
Sample : PB148097ZHE#15
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB148097ZHE#15

Quant Time: Oct 12 04:01:22 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100522W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 06 05:37:50 2022
Response via : Initial Calibration

