

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101123\
 Data File : VN079516.D
 Acq On : 11 Oct 2023 14:48
 Operator : JC\MD
 Sample : 04819-05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TB-01-100923

Quant Time: Oct 12 06:09:40 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100523W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 06 03:34:49 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	284642	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	567695	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	538763	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	180580	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	230236	49.886	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	99.780%
35) Dibromofluoromethane	8.171	113	202831	49.718	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.440%
50) Toluene-d8	10.571	98	664500	45.016	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	90.040%
62) 4-Bromofluorobenzene	12.853	95	232296	49.036	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.080%
Target Compounds						
95) Naphthalene	15.647	128	16245	2.695	ug/l	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101123\
Data File : VN079516.D
Acq On : 11 Oct 2023 14:48
Operator : JC\MD
Sample : 04819-05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TB-01-100923

Quant Time: Oct 12 06:09:40 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100523W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 06 03:34:49 2023
Response via : Initial Calibration

