

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101123\
 Data File : VN079518.D
 Acq On : 11 Oct 2023 15:36
 Operator : JC\MD
 Sample : 04819-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SW-AP-01-100923-FD

Quant Time: Oct 12 06:10:23 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100523W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 06 03:34:49 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	282988	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	545398	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	509947	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	162245	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	223042	48.610	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	97.220%
35) Dibromofluoromethane	8.171	113	198855	50.736	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.480%
50) Toluene-d8	10.571	98	639531	45.096	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	90.200%
62) 4-Bromofluorobenzene	12.853	95	211567	46.486	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.980%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101123\
Data File : VN079518.D
Acq On : 11 Oct 2023 15:36
Operator : JC\MD
Sample : 04819-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SW-AP-01-100923-FD

Quant Time: Oct 12 06:10:23 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100523W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 06 03:34:49 2023
Response via : Initial Calibration

