

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101123\
 Data File : VN079526.D
 Acq On : 11 Oct 2023 18:48
 Operator : JC\MD
 Sample : 04833-06
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 GW-11-23-26

Quant Time: Oct 12 06:13:59 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100523W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 06 03:34:49 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	294545	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	572958	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	543384	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	166584	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	236925	49.609	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	99.220%
35) Dibromofluoromethane	8.171	113	207255	50.335	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.680%
50) Toluene-d8	10.571	98	667556	44.808	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	89.620%
62) 4-Bromofluorobenzene	12.853	95	223653	46.778	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.560%
Target Compounds						
64) Tetrachloroethene	11.106	164	1998	0.535	ug/l	Qvalue # 64

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101123\
Data File : VN079526.D
Acq On : 11 Oct 2023 18:48
Operator : JC\MD
Sample : 04833-06
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
GW-11-23-26

Quant Time: Oct 12 06:13:59 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100523W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 06 03:34:49 2023
Response via : Initial Calibration

