

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101123\
 Data File : VN079530.D
 Acq On : 11 Oct 2023 20:24
 Operator : JC\MD
 Sample : 04841-03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TB-7

Quant Time: Oct 12 06:15:25 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100523W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 06 03:34:49 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	278338	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	549486	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	517590	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	164099	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	225927	50.061	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	100.120%
35) Dibromofluoromethane	8.171	113	198186	50.189	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.380%
50) Toluene-d8	10.571	98	642336	44.957	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	89.920%
62) 4-Bromofluorobenzene	12.853	95	215866	47.078	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.160%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101123\
Data File : VN079530.D
Acq On : 11 Oct 2023 20:24
Operator : JC\MD
Sample : 04841-03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TB-7

Quant Time: Oct 12 06:15:25 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100523W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 06 03:34:49 2023
Response via : Initial Calibration

