

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101124\
 Data File : VN084396.D
 Acq On : 11 Oct 2024 12:07
 Operator : JC\MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Oct 14 02:06:10 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 01 07:11:01 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	157662	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	270821	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	231452	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	89729	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	113141	48.400	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	96.800%
35) Dibromofluoromethane	8.165	113	88662	49.659	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.320%
50) Toluene-d8	10.565	98	325900	49.614	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.220%
62) 4-Bromofluorobenzene	12.847	95	108976	45.538	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	91.080%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101124\
Data File : VN084396.D
Acq On : 11 Oct 2024 12:07
Operator : JC\MD
Sample : IBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Oct 14 02:06:10 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 01 07:11:01 2024
Response via : Initial Calibration

