

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN101218\
 Data File : VN051829.D
 Acq On : 12 Oct 2018 13:38
 Operator : MD\SY
 Sample : VN1012WBL01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1012WBL01

Quant Time: Oct 12 16:22:23 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\624N101218W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Oct 12 10:23:44 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	7.20	128	123999	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	8.59	114	699565	30.00	ug/l	0.00
57) Chlorobenzene-d5	11.41	117	559505	30.00	ug/l	0.00

System Monitoring Compounds

27) 1,2-Dichloroethane-d4	8.03	65	234810	31.12	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	103.73%
60) 4-Bromofluorobenzene	12.40	95	189407	23.48	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	78.27%
63) Toluene-d8	10.09	98	789201	32.48	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	108.27%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN101218\
Data File : VN051829.D
Acq On : 12 Oct 2018 13:38
Operator : MD\SY
Sample : VN1012WBL01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
VN1012WBL01

Quant Time: Oct 12 16:22:23 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\624N101218W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Fri Oct 12 10:23:44 2018
Response via : Initial Calibration

