

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN101218\
 Data File : VN051847.D
 Acq On : 12 Oct 2018 21:59
 Operator : MD\SY
 Sample : PB113779TB
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 28 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 PB113779TB

Quant Time: Oct 13 02:28:34 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N101118W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 12 09:51:20 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	839589	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1207178	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1046506	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	365097	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	461557	57.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	115.04%	
35) Dibromofluoromethane	7.59	113	448597	59.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	119.52%	
50) Toluene-d8	10.09	98	1649317	58.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	117.92%	
62) 4-Bromofluorobenzene	12.40	95	465484	48.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.42%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.82	43	17622	10.18	ug/l	99
20) Methylene Chloride	4.56	84	13324	1.86	ug/l #	81
43) Isopropyl Acetate	8.17	43	217871	21.44	ug/l #	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN101218\
Data File : VN051847.D
Acq On : 12 Oct 2018 21:59
Operator : MD\SY
Sample : PB113779TB
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 28 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
PB113779TB

Quant Time: Oct 13 02:28:34 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N101118W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 12 09:51:20 2018
Response via : Initial Calibration

