

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN101218\
 Data File : VN051852.D
 Acq On : 13 Oct 2018 00:02
 Operator : MD\SY
 Sample : PB113779ZHE#05
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 33 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 PB113779ZHE#05

Quant Time: Oct 13 03:41:03 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N101118W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 12 09:51:20 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	813458	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1175185	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1038809	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	355047	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	442387	56.90	ug/l	0.00
Spiked Amount	50.000		Recovery	=	113.80%	
35) Dibromofluoromethane	7.59	113	440510	60.28	ug/l	0.00
Spiked Amount	50.000		Recovery	=	120.56%	
50) Toluene-d8	10.09	98	1628100	59.79	ug/l	0.00
Spiked Amount	50.000		Recovery	=	119.58%	
62) 4-Bromofluorobenzene	12.40	95	457795	48.70	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.40%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.82	43	18374	10.96	ug/l	98
18) Methyl Acetate	4.34	43	9128	1.01	ug/l	94
20) Methylene Chloride	4.55	84	11813	1.70	ug/l #	84
43) Isopropyl Acetate	8.17	43	266779	27.12	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN101218\
Data File : VN051852.D
Acq On : 13 Oct 2018 00:02
Operator : MD\SY
Sample : PB113779ZHE#05
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 33 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
PB113779ZHE#05

Quant Time: Oct 13 03:41:03 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N101118W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 12 09:51:20 2018
Response via : Initial Calibration

